

AR 5-402 Change List:

- In General Principles box, updated E.: **SPECIFICATIONS:** to say all equipment must be latex-free (no more “if available”) and throughout the document, removed occasional redundant specifications to latex-free. Also amended “Disposable is preferred,” adding, “unless otherwise specified,” because for several items, only disposable is allowed (it was termed “single-use” – now all changed to disposable only, for consistency). Same as changes in AR 5-401
- 1.1 – Modified supraglottic airway size language:
 - Original: Sizes covering all patients, from infant to large adult
 - Updated: Sizes covering all patients and weight ranges
- 3.8 – added language “manufacturer recommended” to defibrillator pads
- 4.4 – added SMALL size to CPAP in order to match BLS CPAP requirement in AR 5-401. Also added, “High pressure outlets needed at oxygen sources, if required by device,” as added in AR 5-401
- 4.7, 4.8, 4.9 – removed glucometer, test strips & lancets, this is duplicated on AR 5-401.
- New 4.9 – added Portable CMED radio meeting the requirements of the statewide EMS communications plan